

# SAVI™

# Instructions for Use

## SAVI Patient Manual



Caution: Federal law (U.S.) restricts this device to sale by or on the order of a physician.

# Table of Contents

Before You Begin .....	3
Intended Use .....	4
Warnings and Precautions .....	4
Contraindications .....	6
Clinical Trial and Adverse Reactions .....	7
Your SAVI .....	8
Getting to Know Your Device .....	9
Using Your SAVI .....	11
Setting Up the Device .....	11
Recommended Treatment .....	11
Preparing for Treatment .....	12
Positioning the Device .....	14
Delivering the Treatment .....	15
Renewing Your Prescription by Remote Activation .....	15
Renewing Your Prescription by optional Micro SIM Rx card .....	15
Replacing the Micro SIM Rx card .....	16
Additional Information .....	17
Traveling with your SAVI .....	18
Caring For Your SAVI .....	18
Understanding System Display Messages .....	19
Troubleshooting .....	20
Service .....	21
Return Goods Policy .....	22
Technical Specifications .....	22
Operating Environment .....	22
Storage Environment .....	23
Industry Standard Classification .....	23
REACH and Warning Statement .....	23
EMC Compliance and Warning Statement .....	24
EMC Guidance .....	24
Glossary of Abbreviations .....	29
Medical Device Reporting .....	30
Warranty and Limitation of Liability .....	30
Customer Care Contact Information .....	30

## Before You Begin

Thank you for choosing SAVI™ from eNeura. Please call Customer Care if you do not understand something in this manual.

eNeura and your doctor are committed to helping you manage your migraine pain. The SAVI is a new, safe, non-drug therapy to treat migraine at the first sign of pain.

Please read this entire manual before using the SAVI. Learn the contraindications, cautions, warnings and notes about the use of the device. As the manufacturer, eNeura cannot and does not intend to give medical advice. Contact your doctor for all medical advice. This manual should be kept near the SAVI and be available at all times.

eNeura is committed to the service and support of our customers. If there are any questions about the use of the eNeura SAVI, please contact Customer Care or your local representative at the following:



**Manufactured by:** eNeura Inc.  
101 W. Dickman Street, Ste 900

**eNeura Inc.**  
Corporate Headquarters  
101 W. Dickman Street, Ste 900  
Baltimore, MD 21230 USA  
Tel: +1 408.213.3136  
Toll free (USA only): +1 833.499.9300  
Fax: +1 877.264.1818  
[www.eNeura.com](http://www.eNeura.com)  
[info@eNeura.com](mailto:info@eNeura.com)  
[customercare@eNeura.com](mailto:customercare@eNeura.com)

eNeura (UK) Ltd.  
6th Floor  
One London Wall London EC2Y 5EB  
United Kingdom  
[www.eNeura.co.uk](http://www.eNeura.co.uk)  
Tel. +44 (0) 20.3356 4889  
Fax: +44 (0) 20.7785.8152  
[customercare@eneura.co.uk](mailto:customercare@eneura.co.uk)

### Authorized Representative

Emergo Europe BV  
Prinsessegracht 20  
2514 AP The Hague  
The Netherlands  
Tel: +31 (0) 70.345.8570  
Fax: +31 (0) 70.346.7299  
[www.emergoeurope.eu](http://www.emergoeurope.eu)

### Australia Sponsor

Emergo Australia  
Level 20, Tower II, Darling Park  
201 Sussex Street  
Sydney, NSW 2000  
Australia  
[www.emergogroup.com](http://www.emergogroup.com)

## Intended Use

The SAVI (The System) is indicated for the acute and prophylactic treatment of migraine headache in adolescents (age 12 and older) and adults.

The System is designed for self-treatment and delivers a non-invasive, brief, single pulse of magnetic energy to the back of the head. This creates a brief electrical current in the brain intended to stop or reduce the effects of migraine headaches.

The System is a drug-free treatment option that can be used in the home or away from home based on your doctor's instructions. After treatment, there are no restrictions. You can resume your normal activities.

- ▲ **WARNING:** This device should be used under the supervision of a physician.
- ▲ Keep the SAVI out of the reach of children.
- ▲ Safety and effectiveness have not been established in pregnant women and children under the age of 12.
- ▲ The long-term effects of single-pulse transcranial magnetic stimulation are unknown.

## Warnings and Precautions

The words ▲ **WARNING**, **Precaution** and **NOTE** have special meanings in this manual. Read them throughout the manual to ensure the safe and effective use of your SAVI.

▲ **WARNING:** A ▲ **WARNING** tells you that the personal safety of the patient may be involved. Ignoring a **WARNING** could result in injury to the patient. **WARNINGS** in the manual are shown in an orange box.

**Precaution:** A **Precaution** means that exact steps must be followed to prevent damage to the product. **Precautions** in the manual are shown in a purple box.

**NOTE:** A **NOTE** gives special information to ease product use or to explain important information. **NOTES** in the manual are shown in a dashed box.

- ▲ **WARNING:** The SAVI should be used under the continued supervision of a physician. The System has been prescribed by your physician only for you.
- ▲ Inspect the System for any signs of damage before use. Do not use it if it is cracked or wet. If you suspect damage to the device, call eNeura at: +1 833.499.9300 option 1 for assistance.
- ▲ Do not operate the System in or near an area where explosive gases are being used or have been used. Do not operate near gasoline or natural gas.
- ▲ Do not operate the System in or near the presence of a FLAMMABLE ANESTHETIC MIXTURE WITH AIR or WITH OXYGEN or NITROUS OXIDE.
- ▲ Risk of electrical shock. Do not open the System. There are no parts that can be serviced or replaced by the user. High voltage may be present.
- ▲ Risk of electrical shock. Do not allow the System or power cords to get wet. Quickly wipe up spills on or near the SAVI. Do not use the System in or near water. For example, do not use while in the bathtub or shower, in the rain, or while standing in water or on a wet surface.
- ▲ Do not use the System if the cause of your headache is illness, underlying pathology, trauma or overuse of medication. See your physician if you are uncertain.
- ▲ Do not use the System if you have suspected or diagnosed epilepsy or a personal or family history of seizures. Consult your physician before using the System if a family member has epilepsy or seizures or if you have had seizures, a head trauma or head injury or take any medication such as tricyclic antidepressants, neuroleptic agents, or other drugs that lower the seizure threshold.
- ▲ Do not use the System if you have a history of stroke.
- ▲ The System has not been demonstrated as safe or effective when treating cluster headache.
- ▲ Do not use the System if you use a wearable cardioverter defibrillator (WCD).
- ▲ The long-term effects of chronic magnetic stimulation are unknown.
- ▲ Transcranial magnetic stimulation should only be applied to the back of the head as described in the "Using Your Device" section of this manual.
- ▲ Do not stimulate over the front of the neck or mouth. Severe spasm of the laryngeal and pharyngeal muscles may occur, and the contractions may be strong enough to close the airway or cause difficulty in breathing.
- ▲ Do not stimulate over the upper side of the neck. Stimulation of the carotid sinus nerves, particularly in patients with a known sensitivity to the carotid sinus reflex, could result in a sudden drop in blood pressure, slowing of the heart or loss of consciousness.
- ▲ Do not stimulate the chest or back. The induction of electrical current into the heart may cause cardiac arrhythmias.

## Contraindications

▲ **WARNING:** Failure to follow the restrictions listed below could result in serious injury or death.

The SAVI creates a very strong single-pulse magnetic field. The System has been prescribed by your doctor for your use only. The System may not be used in patients who have metals, conductive materials, or metal-containing implants in their head, neck or upper body. Metals and conductive materials can be affected by a magnetic field. You should discuss this with your doctor before use.

Do not use the System if you have a cardiac pacemaker, vagus stimulator (VNS) or other implanted neurostimulator, implanted cardioverter defibrillator (ICD) or any implanted medical device that stimulates the body or uses any signal from the body. Talk to your doctor before using the System if you receive an implant. Patients with implants affected by a magnetic field should not use the System. Examples of such implants include:

- Aneurysm clips or coils
- Cochlear implants
- Cerebral spinal fluid shunts
- Bullets or pellets lodged in the head or upper body
- Metal plates, screws, staples or sutures in skull, neck, shoulders, arms or hands
- Facial tattoos with metallic ink
- Radioactive seeds
- Metallic artificial heart valves
- Stents
- Filters
- Magnetically programmable shunt valves
- Electrodes

Dental implants, fillings or other dental appliances are okay and are not affected by the device.

**NOTE: If you have an implanted medical device but are not sure if it has metals or conductive materials, please talk to your doctor before using the SAVI.**

**WARNING:** The System should not be used while driving, operating machinery or during any activity in which involuntary muscle contractions may put the user at risk of injury.

- ▲ Stay at least **2 feet (0.6 meter)** from others when using the System. The System could be harmful to anyone with an electronic implant such as a pacemaker. Anyone with a hearing aid or cochlear implant may hear an audible click.
- ▲ The device could be disrupted by RF-emitting equipment including: wireless home network devices, mobile phones, cordless telephones and their base stations and walkie-talkies. See “EMC Compliance and Warning Statement” section for additional information on preventing unwanted interference.

- **Precaution:** Keep the System away from other electronic devices that depend-on (receive) or radiate (transmit) radio frequency energy, when it is powered on.
- The operation of the System may be impaired when operated near home devices such as wireless network routers, mobile phones, cordless telephones and their base stations and walkie-talkies. Keep the SAVI device at least **2 feet (0.6 meter)** from these devices when it is powered on and in use.
- Keep credit cards, audio/video tapes, computers, computer disks, flash memory sticks, cell phones, personal digital assistants (PDAs), MP3 players, headphones, digital cameras, portable glucose meters and other electronic devices or electronic storage media more than **2 feet (0.6 meter)** away from the System when it is in use.
- Keep any loose metal objects such as eyeglasses, keys, coins, jewelry, watches and hair clips more than **2 feet (0.6 meter)** away from the System when it is in use.
- Keep wearable medical devices such as insulin pumps, medicinal pumps, monitors, bone growth stimulators and Transcutaneous Electrical Nerve Stimulator (TENS) devices more than **2 feet (0.6 meter)** away from the System when it is in use.
- Safety and effectiveness have not been established in pregnant women and children under the age of 12.
- Caution should be used for patients with suspected or diagnosed heart problems.
- The System is only intended to be serviced or maintained by the manufacturer. Do not attempt to open the device. The warranty may be invalidated. If the device is opened, contact eNeura at: +1 833.499.9300 option 1.
- Keep the System out of the reach of children.
- Side effects can include minor dizziness, nausea, vomiting, application site tenderness, muscle spasm, headache and migraine.
- Special precautions regarding Electromagnetic Compatibility (EMC) are required when installing and using the System. Portable and mobile communications devices can affect proper operation of the System. See “EMC Compliance and Warning Statement” section of the Instructions for Use for more information.

## Clinical Trial and Adverse Reactions

eNeura completed a clinical study in Headache Centers in the United States designed to support safety and effectiveness of the device for the acute treatment and prophylaxis of migraine headache. Baseline medication and symptoms were recorded for 28 days via patient diary. Subjects were instructed to use the device daily (morning and evening) and for the acute treatment of attacks for three months with no change in preventive medication. Study results showed statistically significant reduction in migraine headache days of 2.8 days from a baseline mean of 9.1 days. Forty-six percent of the patients saw at least a 50% reduction in headache days (responder rate). Reduction in acute medication was 2.9 days. A statistically significant improvement in quality of life was reported. Six patients reported using prophylactic medications at baseline (2.8%), of these, 5 were using topiramate and propranolol.

The efficacy of the eNeura device for prophylactic treatment is based on the result of an open label study. In open label studies, bias may affect the result. Additionally, open-label studies may introduce placebo rates of 10-25%. This is consistent with placebo rates reported in Randomized Controlled studies for migraine prevention.\*

For further information about the clinical trial and any adverse reaction, please contact eNeura at 1-855-366-8355, option 1 or consult your prescribing physician


\*Macedo A, Banos JE, Farre M. Placebo response in the prophylaxis of migraine: a meta-analysis. Eur J Pain 2008; 12;68-75

## Your SAVI

The complete SAVI includes this manual and the following items:

---


The battery-powered, rechargeable device



---


Carrying bag for SAVI and its AC adapter battery charger

**Note:** The carrying bag is part of the packaging for shipping the SAVI device. Please place the device into the bag prior to boxing the device for return/exchange..



---


Battery Charger 12V DC 1.5A 18 watts  
(reorder no. DWG-0505)



---

**NOTE: Do not discard the box or packing materials. They will be required if you need to return this product.**

## Getting to Know Your Device



**A. Power Button**  In the center of the panel on the top of the System. Press the power button to turn the device on and off.

**B. Power Indicator**  LED light on the top of the System inside the Power button. Static white illumination shows the device is on and ready. Pulsing blue illumination indicates device is connecting to the cellular network to transfer data.

**C. Treatment Progress Indicator**  LED light around the power button on the top of the System, shows status as it prepares for treatment.

**D. Prescription Status Indicator**  Located on left side on the top of the System. Confirms a valid prescription is available and shows the status of the prescription: White illumination indicates Rx is active; static orange indicates Rx will expire soon; static orange with orange Power Button indicates Rx has expired.

**E. Contact eNeura Customer Care**  Action required. Call Customer Care.

**F. Temperature warning**  The device temperature is not in range for safe use.

**G. Lock Indicator**  On the right of the Power button. Indicator is visible when device is turned on if the security lock switch is enabled.


**H. AC Adapter**  On the right of the Power button. Indicator is lit when the AC adapter is connected to the device.

**I. Battery Capacity**  On right side on the top of the System. Indicates whether or not battery power is enough to allow treatment. White illumination indicates battery power is available; orange illumination indicates battery power will soon be or has been exhausted.

**J. Treatment Buttons** On the right and left edge of the panel on the top of the device. Press one or both buttons to deliver a treatment.

Read “Understanding System Display Messages” for more information.

## Battery Charger (Power Supply) and Micro SIM Rx card Ports




Door side of device, with door open

**K. Lock ON/OFF Switch** – Located behind accessory port door, on the left. Slide the switch to the left or right to unlock or lock the device.

**L. Volume ON/OFF Switch** – Located behind accessory port door, on the right. Slide the left or right to turn the sound on or off

**M. Battery Charger Port** – Located behind accessory port door, in the center. Recharge the batteries by plugging the AC battery charger, DWG-0505, into this port

AC battery charger 12V DC  
1.5A 18 watts  
(reorder no. DWG-0505)



**Plug that connects to the System**

**N. Micro SIM Rx card port** – Located on the back of the System beneath the Accessory Door. Insert the micro SIM RX card into this port.

- **Precaution:** Use only eNeura-supplied accessories with your System: Battery charger (reorder no. DWG-0505) and Micro SIM Rx card.

## Using Your SAVI

- **Precaution:** Prior to use of the System, this manual must be completely read and understood. Improper use may cause personal injury and damage to the System and may void the warranty.
- You, as the user of the System, must become familiar with the functions and features of the device prior to use.

## Setting Up the Device

Be prepared to treat your migraine when needed. Set up your SAVI as soon as it arrives and keep your device charged and ready to treat.

**NOTE: Use only as directed. The device should only be used per the instructions of your physician who prescribed the System.**

- Remove the System and the battery charger from the box.

**NOTE: Do not discard box or packing materials. They will be required if you need to return this product.**

- Your device should arrive ready to use. If it is not ready, plug the battery charger into a standard wall outlet. Connect the other end of the battery charger to the port on the device. Charge the battery for about 4 to 6 hours.
- Treatment delivery is not available during battery charging.
- If the device is on while charging, the AC adapter symbol and the battery capacity symbol will be lit.
- Unplug the battery charger when the device is fully charged. A fully charged battery should deliver at least 12 treatments.
- Keep the accessory door closed when not in use.
- Store the device and its battery charger in a cool, dry place, away from excessive dust and direct sunlight.

- **Precaution:** Only use the battery charger (DWG-0505) included with the SAVI. Contact eNeura if you have any questions.

## Recommended Treatment

The treatment with the SAVI should be performed per the instructions of your prescribing physician.

### Prevention

Treat with 4 pulses each morning and evening  
2 consecutive pulses  
Wait 15 minutes  
Repeat the 2 consecutive pulses

### Acute

3 sequential pulses (early) at the onset of migraine pain  
Wait 15 minutes  
If needed treat with additional 3 pulses  
Wait 15 minutes  
If needed treat with additional 3 pulses

Read the instructions below prior to operating the device. Then, follow the entire procedure as described below. Consult your physician for additional treatment information.

**NOTE: Become familiar with the System and how to use it before you experience a migraine.**

## Preparing for Treatment

Consult your doctor about your specific condition and the best timing of your sTMS treatment.

- **Precaution:** Keep the System away from other electronic devices that depend-on (receive) or radiate (transmit) radio frequency energy, when it is powered on.
- The operation of the System may be impaired when operated near home devices such as wireless network routers, mobile phones, cordless telephones and their base stations and walkie-talkies. Keep the SAVI device at least **2 feet (0.6 meter)** from these devices when it is powered on and in use.
- The System emits a strong magnetic pulse that may interfere with the operation of common home electronic devices such as radios, televisions, wireless network routers, mobile phones, cordless telephones and their base stations and walkie-talkies, if they are not installed and used in accordance with the manufacturer's instructions. There is no guarantee that interference will not occur in a particular installation. If the System causes interference with other devices, try to correct the matter by:
  - Reorienting or relocating the device receiving the interference.
  - Increase the separation distance between the System and the device.
  - Connect the System to an AC outlet on a circuit different from that to which the other device is connected.

Use the System in a comfortable, seated position. Remove the System from its storage location.

- ▲ **WARNING:** Inspect the System for any signs of damage before use. Do not use if it is cracked or wet. If you suspect damage to the device, call eNeura at +1 833.499.9300 option 1 for assistance.

The System needs to be in direct contact with the back of your head to work properly. Before treatment, remove any hat, head covering, or hairpiece that covers the back of your head and undo any braids, ponytails or buns.

- ▲ **WARNING:** Remove any personal medical devices such as insulin pumps or other medical pumps and bone growth stimulators.

- Press the power button



- Watch the LED track light progress around the power button:



- It takes 30-60 seconds to prepare the System for treatment.
- As the device prepares for treatment, each segment will light up to indicate progress.
- When the system is ready, the device lights up all 6 segments and sounds an audio tone:




- Once the System is ready, you have 45 seconds to position the device and deliver treatment.

- ▲ **WARNING:** Stay at least **2 feet (0.6 meter)** from others when using the SAVI. This device could be harmful to anyone with an electronic implant such as a pacemaker.
- ▲ Remove any personal medical devices such as insulin pumps, other medicinal pumps, bone growth stimulators, Transcutaneous Electrical Nerve Stimulator (TENS) devices and hearing aids.
- ▲ Do not use the System if you use a wearable cardioverter defibrillator (WCD).
- ▲ The System should not be used while driving, operating machinery or during any activity in which involuntary muscle contractions may put you at risk of injury.
- ▲ Risk of electrical shock: Do not allow the System or power cords to get wet. Quickly wipe up spills on or near the device. Do not use the device in or near water. For example, do not use while in the bathtub or shower, in the rain or while standing in water or on a wet surface.
- ▲ Do not operate the System in or near an area where explosive gases are being used or have been used. Do not operate near diesel fuel, gasoline or natural gas.
- ▲ Do not operate the System in or near the presence of a **FLAMMABLE ANESTHETIC MIXTURE WITH AIR** or with **OXYGEN** or **NITROUS OXIDE**.
- ▲ Do not use the System if the cause of your headache is illness, trauma or excess medication. Consult your doctor if you are unsure.
- ▲ Do not use the System if you have suspected or diagnosed epilepsy. Consult your doctor before using the System if a family member has epilepsy. Consult your doctor if you have had a head trauma, head injury or take any medication such as tricyclic antidepressants, neuroleptic agents or other drugs that lower the seizure threshold.
- ▲ Do not use the System if you have a history of stroke.
- ▲ Caution should be used for patients with suspected or diagnosed heart problems.

- **Precaution:** Keep the SAVI away from metal or conductive objects, medical devices and magnetic media when it is powered on.
- When activated, the System emits a strong magnetic pulse and may interfere with other metal, electronic or magnetic products within **2 feet (0.6 meter)**. A person with a hearing aid or cochlear implant may hear a click when the System is activated.
- The operation of the SAVI System may be impaired when operated near home devices such as wireless network routers, mobile phones, cordless tele- phones and their base stations, and walkie-talkies. Keep the SAVI device at least **2 feet (0.6 meter)** from these devices when it is powered on and being used.
  - Move credit cards, audio/video tapes, computers, computer disks, flash memory sticks, cell phones, personal digital assistants (PDAs), MP3 players, headphones, digital cameras, portable glucose meters and other electronic devices or electronic storage media more than **2 feet (0.6 meter)** away from the System.
- Remove all metal objects from your head, face, neck, arms and hands. Remove eyeglasses, hearing aids, removable metallic orthodontic appliances, hair clips and earphones and move them more than **2 feet (0.6 meter)** away from the System.
- Remove any loose metal objects from your shirt pockets such as keys, coins, clips, pens and mechanical pencils and move them more than **2 feet (0.6 meter)** away from the System.

## Positioning the Device

- ❑ With the SAVI in front of you: The treatment buttons should be facing up and the accessory door should be facing away from you. Place your hands on the sides of the device with your thumbs on the treatment buttons.
- ❑ Lift the System over your head. The device should be placed so that it comfortably and naturally cradles the base of your skull. Hold it firmly against your head.
- ❑ The curved surface positions the device to fit comfortably against the head and optimize pulse delivery.



- ▲ **WARNING:** Transcranial magnetic stimulation should only be applied to the back of the head as described in the “Positioning the Device” section of this manual.
- ▲ Do not stimulate over the front of the neck or mouth. Severe spasm of the laryngeal and pharyngeal muscles may occur. Contractions may be strong enough to close the airway or cause difficulty breathing.
- ▲ Do not stimulate over the side of the neck. Stimulation of the carotid sinus nerves, particularly in patients with a known sensitivity to the carotid sinus reflex, could result in a sudden drop in blood pressure, slowing of the heart or loss of consciousness.
- ▲ Do not stimulate the chest or back. Electrical current into the heart may cause cardiac arrhythmias.

**NOTE:** If the device detects an error when it is on, a symbol is displayed. See the “Understanding System Display Messages” and “Troubleshooting” sections of this manual.

**NOTE:** The System is a handheld device that comes in direct contact with skin. The housing of the device is made of polycarbonate, a plastic common in consumer medical products, and poses no handling risk.

## Delivering the Treatment

**NOTE: Two hands are needed to stabilize the device, but only one treatment button needs to be activated to deliver treatment.**

**NOTE: The device shuts itself off if treatment is not delivered within 45 seconds. Press the power button to turn the device back on and restart the treatment procedure.**

- Once the device is in place and your thumbs are on the treatment switches, press one or both switches and hold for at least 2 seconds.
- The device emits a soft, audible click as the treatment is delivered.
- After you have successfully delivered the first pulse, remove the device from your head.
- The LED light on the POWER button softly pulses to show that treatment is completed. The device is now ready for another pulse. You may select to treat again by pressing the power button.
- Repeat the subsequent procedures from page 15, Positioning the Device.
- After delivering the final pulse and 30 seconds of inactivity, the device will attempt to connect to a cellular network, transfer data and will then power off.
- While the power button is lit **Blue** place the device on a flat surface at least 8 inches (20 cm) away from you, other people and electronic equipment.
- Delivery of several pulses in a row may cause the device's surface temperature to exceed 105.8° F (41° C). The surface may get as warm as 118.4° F (48° C). Let the device cool for 15 minutes. Try again.

## Renewing Your Prescription by Remote Activation

**NOTE: The System is intended for use under the care of a doctor. For more information on renewing your prescription, please contact eNeura Customer Care. An active prescription is required.**

**NOTE: The System only delivers sTMS treatments if the programmed prescription has not expired.**

- The device will display an ORANGE warning light when prescription is about to expire:  Renew your prescription promptly.
- Contact Customer Care to request a refill.
- Upon receipt of payment, Customer Care will initiate reactivation of your device via the cellular network.
- Ensure the System lock switch is in the OFF position and the AC adapter is not connected to your device.
- Press the Power Button on your device. After approximately 30 seconds, pulsing blue illumination on the Power Button  indicates the device is attempting to connect to a cellular network.
- While the power button is lit **Blue** place the device on a flat surface at least 8 inches (20 cm) away from you, other people and electronic equipment.
- After the device has successfully connected to a network, the new Rx data will transfer, and the ORANGE Rx warning light will become white again. You may resume use.

## Renewing Your Prescription by OPTIONAL Micro SIM Rx card

- The Micro SIM Rx card from your prescription card may be installed to activate the device to deliver sTMS treatments.
- The device will display an ORANGE  warning light when your Micro SIM Rx card is about to expire: Renew your prescription promptly.
- After renewing your prescription, you will receive a prescription card containing a new Micro SIM Rx card within 2-3 business days.

## Replacing the Micro SIM Rx card

- ❑ Open the accessory door on the back of the device.



- ❑ Push on the edge of the Micro SIM Rx card. It pops out about 3 mm (1/8 inch).
- ❑ Pull the Micro SIM Rx card out with your fingers. This expired Micro SIM Rx card is of no value.
- ❑ Dispose as e-waste or recycle by appropriate methods.

▲ **WARNING:** The Micro SIM Rx card may be swallowed and present a choking hazard for infants or small children if it is ejected from the SIM card port and left loose and unattended.

- ❑ Remove your new Micro SIM Rx card (the cut-out portion) from the prescription card by pressing down firmly on the Micro SIM Rx card. Check that it is the same shape as shown below. Remove any excess plastic.

Micro SIM Rx card



- ❑ Insert the Micro SIM Rx card with the gold side up and the edge with the blunt corner pointing into the Micro SIM Rx card port
- ❑ Push the Micro SIM Rx card until it is flush with the opening and snaps into place. Close the accessory door.



## Additional Information


### Battery Capacity Indicator



Your SAVI has rechargeable batteries. The battery capacity indicator on your device shows how much charge is remaining in the battery. eNeura recommends that you keep your device's battery charged so that it is always ready when you need it. Typically, a fully charged battery can deliver 18 treatments. Your battery pack life is roughly a minimum of 100 charge cycles.

### Recharging the Battery

- ❑ To recharge your battery, push the round plug on the battery charger into the battery charger port on the back of the device. Plug the battery charger into a standard wall outlet.



**Connect AC adapter here**

- ❑ The AC adapter must be disconnected to allow the device to connect to a cellular network. Wait for the device to shut itself off before connecting the AC adapter.
- ❑ The AC adapter, power button and battery power indicator symbols will be lit for approximately 10 seconds when the AC adapter is connected to the device. After 10 seconds, the lights will shut off while the AC adapter continues to charge the battery pack.
- ❑ Treatment while recharging is disabled for safety. To use the device for treatment, unplug the battery charger and proceed with your treatment.

**▲ WARNING:** The battery charger cable may present a strangulation or trip hazard if the System is left unattended around infants and small children. Recharge the System's battery in a safe location, away from foot traffic and children. Store the battery charger in a safe place when not in use.

### Locking the System/Airplane Mode

- ❑ The System lock switch is located beneath the Accessory door.
- ❑ The System lock may be controlled by sliding the switch to the desired position (ON or OFF).
- ❑ When the System is locked, the device will prevent others from using it by disabling treatment functionality.
- ❑ The System lock switch must be OFF to allow the device to connect to a cellular network. Wait for the device to shut itself off before you lock it.
- ❑ Activating the System lock switch places the device in Airplane Mode. Place the System in Airplane Mode whenever you are in a location where use of cellular devices is discouraged or forbidden.

**Lock ON**



## Sound OFF



### Adjusting the Audible Tone

- The System sounds an audible tone when the device is ready to deliver a treatment.
- The switch to control the audible tone is located beneath the Accessory door.
- Slide the switch to the desired position (ON or OFF).

## Traveling with your SAVI

- Your SAVI may travel with you. Place the device in Airplane Mode as explained on page 18 (see “Locking the System/Airplane Mode”) and pack it securely to avoid damage during travel. Before using the unit after traveling, check it for any damage to ensure it appears to be working correctly before you use it.
- Whenever possible, pack your SAVI in a carry-on rather than in checked luggage.
- At security, remove your SAVI from your carry-on, place it in a bin and, before it goes through x-ray, notify TSA or Airport Security that it is a medical device. This will speed up the security-check process.
  - Security may wipe down the device with a chemically sensitive wipe, which is a standard security procedure and will not harm the device.
- For international travel, use the correct universal travel power adapter. A power converter is not needed. Only use the AC adapter that was provided with the System.

## Caring For Your SAVI

### Inspect your device regularly

- Before using your SAVI always check to make sure your device is in good working order. If you notice damage to the exterior of the device or rattling when you shake the device DO NOT USE the device and contact your eNeura representative for a replacement.

### Rechargeable Battery

- The rechargeable battery in the SAVI is in a sealed compartment and may only be replaced by a trained service technician. If the battery fails to recharge, please contact Customer Care to arrange for a device exchange.

### Cleaning Procedure

- eNeura recommends cleaning the SAVI every three months, similar to the way you would clean a telephone handset. It may be necessary to clean the System more often if it becomes dirty or sticky. It may be cleaned as many times as needed.
- Only the external surfaces of the System may be cleaned. Wipe dust and lint away with a dry cloth. Take care not to drip or spray liquid directly onto your device or get liquid into any of the ports, the buttons or the display. Dampen a soft cloth in water premixed with a mild detergent or an alcohol-based, hospital-grade, hand-disinfecting solution. Wring out any excess liquid, and then wipe the device. Immediately dry the device with a soft dry cloth to make sure that no liquid remains on the surface.
- Do not use any spray cleaners because fluid should not enter openings. Do not use detergents or other cleaning solutions to clean the display.

- Precaution:** Disconnect the SAVI from the battery charger prior to cleaning. NEVER immerse the System. Do NOT drip or splash any liquid on the device. Contact eNeura if the device gets wet. Do not attempt to sterilize the System.

## Understanding System Display Messages

If your device displays one of the following images, please take the steps listed to clear the message.



### Battery charge is full



### Battery charge is low

- The rechargeable battery charge is low. There is only enough power to deliver approximately 10 treatments so charge the battery at your earliest convenience.
- Charge the battery for at least 15 minutes. To fully charge the battery, allow the device to charge for about 4 hours.
- If the problem persists, check that only the eNeura battery charger (DWG-0505) is in use. Check that the battery icon on the display is lit to show it is charging.
- If the problem persists, contact eNeura or your local Agent for help.



### Valid treatment card installed

- If the Rx symbol is white, the installed prescription is valid and active.



### Prescription expires in less than 15 days

- If the Rx symbol is orange, please contact eNeura or your local Representative to order a new prescription card. An active prescription is required.



### Invalid treatment card installed

- If the Rx symbol is orange and pulsing, the Micro SIM Rx card installed in the device cannot be read by your device and may be invalid.
- Turn off the device, remove and reinstall the Micro SIM Rx card.
- If the Rx symbol is still orange and pulsing, contact eNeura or your local Representative for assistance.



### Prescription expired

- The prescription installed in the device has expired and can no longer be used. Each prescription is valid for a specific date range based on the prescription issued by your doctor. If you have received a new Micro SIM Rx card, install it in your device.
- If you need a new prescription, contact eNeura or your local Representative for assistance. An active prescription is required. eNeura can contact your physician if a renewal prescription is required.



### Security Lock Enabled

- The security lock is enabled, and the device is non-operable. To unlock the device and make the device operable, slide the lock switch to the "off" position.
- When enabled, the security lock places the System into Airplane Mode and will prevent wireless transmissions from the device.



### Device temperature not in operating range

- ❑ The device does not operate when its temperature is below 16°C (60°F) or above 40°C (104°F). Move the device to a more temperate environment, such as inside your home or office, and wait 15 minutes. Try again.

- **Precaution:** Do NOT actively heat or cool your SAVI in any way.



### Contact eNeura Customer Care

- ❑ If this symbol appears with the Prescription low or expired symbol, contact Customer Care (+1-408-245-6400, option 1) to renew or refill your prescription.
- ❑ Otherwise, this symbol means service is needed and the device will not deliver treatment. Contact Customer Care for assistance.



### Treatment not allowed (Orange power button)

- ❑ If this symbol appears with the Prescription Expired, Device Locked, Battery Low, Charger Connected or other symbols, then the device will prohibit treatment until the other system messages are cleared. Clear these symbols. Restart the device.
- ❑ Contact Customer Care for help if other symbols do not appear with the Orange power button.
- ❑ If treatment is not allowed due to a device failure, an expired or invalid prescription, or temperature out of range condition, then, after 30 seconds of inactivity, the device will attempt to connect to a cellular network, transfer data and power off. During data transfer, the power button indicator will pulse blue.



### Device connecting to a cellular network (Blue power button)

- ❑ This symbol appears when the device is connecting to a cellular network to upload treatment data or download a new Rx. The device will automatically shut down after it has completed the data transfer.
- ❑ While the power button is lit Blue place the device on a flat surface at least 8 inches (20 cm) away from you, other people and electronic equipment.

## Troubleshooting

### None of the device indicator lights are on

- ❑ The device may not be powered on. Press the power button for at least one second.
- ❑ The battery may be completely discharged. Plug in the battery charger to charge the battery for at least 15 minutes. Be sure that only the eNeura battery charger (DWG-0505) is being used. When you plug in the battery charger, the display should light. Check that the battery icon on the display is lit to indicate that the battery is charging.
- ❑ If the problem persists, contact eNeura or your local Representative for assistance.

- ▲ **WARNING:** There are no user serviceable parts inside. Do not attempt to open the device, as it could expose user to hazardous voltages. Do not try to modify the unit in any way.

### The device is charging correctly, but I can't make it deliver a treatment.

- Check the power button indicator  to confirm the device is ready for treatment.
- Check the lock switch to make certain it is not locked.
- Disconnect the battery charger if it is connected to the device.
- The Micro SIM Rx card may not be installed or may have expired. Check the  indicator and install a current Micro SIM Rx card if required.
- The device may be too warm or too cold. Check the  indicator. The device does not operate when its temperature is below 15°C (60°F) or above 40°C (104°F). Move the device to a normal indoor environment and wait 15 minutes before trying again.

- **Precaution:** Do NOT actively heat or cool your SAVI in any way.

### Service

- For questions about your SAVI, please call or email Customer Care at eNeura.
- Please have your device available when calling eNeura Customer Care.
- When the System is no longer usable, return it to eNeura or your local Representative. Do not dispose of the device. For questions, contact eNeura or your local Representative for assistance.

- **Precaution:** Only authorized service personnel should service or repair components of the SAVI. There are no user serviceable parts. Improper service, repair or modifications performed by unauthorized personnel may pose a hazard and void the warranty. Do not heat, incinerate or dispose of the device.

## Return Goods Policy

- ❑ Contact eNeura or your local Representative for information on product returns and return authorization. Contact information is available at the beginning of this manual and on the last page.

**NOTE: Do not discard the box or packing materials. They will be required if you need to return this product.**

## Technical Specifications

Weight:	3.2 lb. (1.4 kg)
Size:	8.8 in long x 5.1 in wide x 2.7 in deep (22.4 cm long x 13 cm wide x 6.9 cm deep)
Electrical Power	Internally powered (Rechargeable battery)
Battery charger:	Input: 100-240V AC, 400 mA (max), 50/60 Hz Output: 12V DC, 1.5A
Plug Type:	Class II (ungrounded 2-prong plug)
Rechargeable Battery:	7.2V, 2150 mAh, 15.48 Wh lithium ion battery Typical battery life is 100 charge cycles (minimum)
Magnetic Pulse Output	0.9 Tesla (nominal) at 1 cm from the center of curved surface of the SAVI per treatment
FCC ID: 2AR4G-HL7648	The SAVI device complies with FCC rules. Highest reported SAR for body-worn condition is 1.056 W/kg and extremities is 1.63W/kg.

## Operating Environment

- ❑ The SAVI is a battery-operated device.
- ❑ The SAVI requires a regular home or office electrical outlet for proper battery recharging.
- ❑ The SAVI operates best in typical household and office environments where the temperature is 15°C to 40°C (60°F to 104°F); relative humidity range of 10% to 90%, non-condensing; an atmospheric pressure range of 700 hPa to 1013 hPa; operating altitude of less than or equal to 3000m.
- ❑ The SAVI is not affected by exposure to dust or lint normally found in household and office environments.
- ❑ The SAVI and its accessories may be exposed to sunlight for a few days duration, but prolonged or repeated exposure to direct sunlight is not recommended.
- ❑ It is recommended that a cold device (below 60°F or 15°C) be allowed to warm to room temperature for at least 15 minutes prior to use. Similarly, it is recommended that a hot device (above 104°F or 40°C) be allowed to cool for at least 15 minutes prior to use.
- ❑ The SAVI is suitable for home use and complies with IEC 60601-1-11:2015.

• **Precaution:** Do NOT actively heat or cool your SAVI in any way.

## Storage Environment

- ❑ The SAVI may be stored at -40°C to 60°C (-40°F to 140°F) with a relative humidity of up to 90% non-condensing and an atmospheric pressure range of 700 hPa to 1060 hPa and storage altitude of less than or equal to 3000m.
- ❑ Keep your SAVI dry during operation, storage and transport. When not in use, store the sTMS to protect it from household dust and lint, accidental spills or exposure to the elements (rain, dust, insects, vermin, prolonged direct sunlight, etc.).
- ❑ During extended storage periods, it is recommended that the battery be periodically recharged every 6 months, so that the device will be available for use when needed.

## Industry Standard Classification

(IEC 60601-1)	Classification
Type of Electric Shock Protection (SAVI) (Battery Charger, external)	Internally powered Class II
Degree of Electric Shock Protection	Type B applied part
Degree of Protection Against Harmful Ingress of Liquids and Particulates	IP22
Safety in the Presence of Flammable Atmosphere	Not suitable for use in flammable environment
Mode of Operation	Short-time operation – individual 1ms pulses
EMC	Class B
Service Life (SAVI device, AC Adapter and Rx SIM Card)	20,000 treatments

## REACH and Warning Statement

Thresholds and Substances of Very High Concern (SVHCs) are defined in Article 7 and 57 of Regulation (EC) No 1907/2006 (REACH Regulation). This article contains a substance on the Candidate List in a concentration above 0.1% (w/w), namely: 4-Nonylphenol, branched and linear -substances with a linear and/or branched alkyl chain with a carbon number of 9 covalently bound in position 4 to phenol, covering also UVCB- and well-defined substances which include any of the individual isomers or a combination thereof. This substance is completely enclosed in the device and does not pose a safety concern during use.

## EMC Compliance and Warning Statement

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

### □ IEC 60601-1-2: 2014

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules and with the limits of the standard for medical devices, IEC 60601-1-2, suitable for use in all environments including domestic. The unit also complies with the requirements of EN 60601-1-2, providing the presumption of compliance to the European Union’s Medical Device Directive 2007/42/EC. The limits are designed to provide reasonable protection against harmful interference in a typical medical installation. This equipment generates, uses, and can radiate radio-frequency energy, (such as wireless home networks, mobile phones, cordless telephones and their base stations, and walkie-talkies), and, if not installed and used in accordance with the manufacturer’s instructions, may cause harmful interference to other devices in the vicinity. Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the SAVI, including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result. However, there is no guarantee that interference will not occur in a particular installation. If this equipment causes interference with other devices, which may be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the device receiving the interference.
- Increase the separation between the equipment.
- Connect the equipment into an outlet on a circuit different from that to which the other device(s) are connected.
- Consult the manufacturer or field service technician for help.

## EMC Guidance

In accordance with Clause 4.3 of IEC 60601-1, the Essential Performance of the sTMS Mini is to not deliver a pulse above 0.95T. EMC interference from other equipment may disrupt normal operation of the sTMS Mini and may impede the delivery of the 0.9T magnetic pulse to the patient’s head. If other equipment interferes with normal sTMS Mini operation, which may be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the sTMS Mini device away from the equipment.
- Increase the separation between the sTMS Mini and the equipment.
- Connect the equipment into an outlet on a circuit different from that to which the sTMS Mini is connected.
- Consult eNeura Customer Service for assistance.

<b>Guidance and Manufacturer’s Declaration - Emissions</b>
The <b>DWG-0600</b> is intended for use in the electromagnetic environment specified below. The customer or user of the <b>DWG-0600</b> should ensure that it is used in such an environment.

Emissions Test	Compliance	Electromagnetic Environment – Guidance
RF Emissions CISPR 11	Group 1	The <b>DWG-0600</b> is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
RF Emissions CISPR 11	Class B	
Harmonics EN 61000-3-2	Class A	
Voltage Fluctuations/Flicker emissions	Complies	

**Guidance and Manufacturer's Declaration – Immunity**

The **DWG-0600** is intended for use in the electromagnetic environment specified below. The customer or user of the **DWG-0600** should ensure that it is used in such an environment.

<b>Immunity Test</b>	<b>IEC 60601 Test Level</b>	<b>Compliance Level</b>	<b>Electromagnetic Environment – Guidance</b>
Electrostatic discharge (ESD) IEC 61000-4-2	±8 kV contact ±15 kV air	±8 kV contact ±15 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines ±1 kV for input/output Lines* <sup>1</sup>	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV common mode	±1 kV differential mode ±2 kV common mode	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	Voltage Dips 30% reduction, 25/30 periods At 0°	Voltage Dips 30% reduction, 25/30 periods At 0°	Mains power quality should be that of a typical commercial or hospital environment. If the user of the EQUIPMENT requires continued operation during power mains interruptions, it is recommended that the EQUIPMENT be powered from an uninterruptible power supply or a battery.
	Voltage Dips > 95% reduction, 0.5 period At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315°	Voltage Dips > 95% reduction, 0.5 period At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315°	
	Voltage Dips > 95% reduction, 1 period At 0°	Voltage Dips > 95% reduction, 1 period At 0°	
	Voltage Interruptions > 95% reduction, 250/300 periods	Voltage Interruptions > 95% reduction, 250/300 periods	
(50/60 Hz) magnetic field IEC 61000-4-8	30 A/m	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

1

The EUT does not contain signal input/output lines longer than 3 m in length.

## EMC Guidance (continued)

### Guidance and Manufacturer's Declaration – Electromagnetic Immunity

The **DWG-0600** is intended for use in the electromagnetic environment specified below. The customer or user of the **DWG-0600** should ensure that it is used in such an environment.

Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment – Guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz (6 Vrms in ISM and amateur radio Bands within 150kHz – 80MHz)	3 Vrms	Portable and mobile RF communications equipment should be used no closer to any part of the EQUIPMENT, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.  <b>Recommended separation distance</b> $d = 1.2\sqrt{P}$
Radiated RF IEC 61000-4-3	10 V/m 80 MHz to 2.7 GHz	10 V/m	$d = 1.2\sqrt{P}$ 80 MHz to 800 MHz $d = 2.3\sqrt{P}$ 800 MHz to 2.7 GHz  where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m).  Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey <sup>a</sup> , should be less than the compliance level in each frequency range. <sup>b</sup>

NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

<sup>a</sup> Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the EQUIPMENT is used exceeds the applicable RF compliance level above, the EQUIPMENT should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the EQUIPMENT.

<sup>b</sup> Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

Immunity to RF Wireless Communications Equipment						
Test Frequency (MHz)	Band <sup>a)</sup> (MHz)	Service <sup>a)</sup>	Modulation <sup>b)</sup>	Maximum Power (W)	Distance (m)	Immunity Test Level (V/m)
385	380 - 390	TETRA 400	Pulse modulation <sup>b)</sup> 18 Hz	1.8	0.3	27
450	430 - 470	GMRS 460, FRS 460	FM <sup>c)</sup> ± 5 kHz deviation 1 kHz sine	2	0.3	28
710 745 780	704 - 787	LTE Band 13, 17	Pulse modulation <sup>b)</sup> 217 Hz	0.2	0.3	9
810 870 930	800 - 960	GSM 800/900, TETRA 800, iDEN 820, CDMA 850, LTE Band 5	Pulse modulation <sup>b)</sup> 18 Hz	2	0.3	28
1720 1845 1970	1700 - 1990	GSM 1800; CDMA 1900; GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS	Pulse modulation <sup>b)</sup> 217 Hz	2	0.3	28
2450	2400 - 2570	Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulation <sup>b)</sup> 217 Hz	2	0.3	28
5240 5500 5785	5100 - 5800	WLAN 802.11a/n	Pulse modulation <sup>b)</sup> 217 Hz	0.2	0.3	9

a) For some services, only the uplink frequencies are included.  
b) The carrier shall be modulated using a 50 % duty cycle square wave signal.  
c) As an alternative to FM modulation, 50 % pulse modulation at 18 Hz may be used because while it does not represent actual modulation, it would be worst case.

**Recommended separation distances between  
portable and mobile RF communications equipment and the sTMS Mini**

The sTMS Mini is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the sTMS Mini can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the sTMS Mini as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitted (m)		
	150 kHz to 80 MHz $d = 1.2\sqrt{P}$	80 MHz to 800 MHz $d = 1.2\sqrt{P}$	800 MHz to 2.7 GHz $d = 2.3\sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance  $d$  in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where  $P$  is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

The device and the method of use that comprise the eNeura SAVI™ are protected by U.S. Patents 6,402,678, 7,294,101, 7,494,458 and 7,601,116; by European Patents 1 307 260 B1 and 1 977 787 B1; and by Canadian Patent 2423840. Additional patents are pending.

## Glossary of Abbreviations

<b>Intertek</b>	Intertek Mark certification marks indicate that this product has been tested to a U.S. standard and it meets the requirements of an applicable ETL standard, or another recognized document used as a basis for certification.
<b>ETL</b>	Conforms to AAMI STD ES60601-1, IEC STDS 60601-1-6, and IEC STD 60601-1-11. Certified to: CSA STD C22.2#60601-1
<b>EMC</b>	Electromagnetic Compatibility
<b>IEC</b>	International Electro-technical Commission
<b>SN</b>	Serial Number
<b>TMS</b>	Transcranial Magnetic Stimulation, which is a method of using a brief pulse of magnetic energy to stimulate nerves in the brain

## Key to Symbols

Intertek Mark	
Protection against falling liquids	
Type BF Applied part	
Caution: Federal law (US) restricts this device to sale by or on the order of a physician	<b>Rx Only</b>
Warning	
Temperature limitation	
Humidity limitation	
Contact eNeura Customer Care	
Ready for Treatment	
Treatment not allowed	
Connecting to cellular network	
AC Adapter connected	

Attention: See Instructions for Use	
Risk of strong magnetic field	
Warning, electricity	
Keep Dry	
Federal Communications Commission Class B – certified for home use	<b>FC</b>
Authorized representative in European Community	
EMC non-ionizing radiation	
Smart Card status	
Power button On/Off	
Preparing for Treatment	
Lock switch enabled	
Rechargeable Battery capacity level	

## Medical Device Reporting

- ❑ Any potential adverse incident involving eNeura products should be reported immediately by calling your local Representative or Customer Care at eNeura.  
Users and physicians may also report adverse events to FDA through MedWatch at 1-800-332-1088 or at <https://www.fda.gov/safety/medwatch>.

## Warranty and Limitation of Liability

- ❑ eNeura Inc. warrants that the product when delivered is free from defect in materials and workmanship and conforms to the manufacturer's product specifications. eNeura shall not be liable for any incidental or consequential loss, damage, or expense, directly or indirectly arising from the use of, or inability to use, its product.
- ❑ THE FOREGOING WARRANTY IS IN LIEU OF AND EXCLUDES ALL OTHER WARRANTIES, WHETHER EXPRESSED OR IMPLIED, ARISING BY OPERATION OF LAW OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.
- ❑ eNeura does not assume, nor authorizes any person to assume for it, any other additional liability or responsibility with respect to this product other than as set forth in writing herein.

## Customer Care Contact Information



**Manufactured by:** eNeura Inc.  
101 W. Dickman Street, Ste 900  
Baltimore, MD 21230 USA

**eNeura Inc.**  
Customer Care Headquarters  
1735 N. First Street, Ste 105  
San Jose, CA 95112 USA  
Tel: +1 833.499.9300  
Fax: +1 877.264.1818  
[www.eNeura.com](http://www.eNeura.com)  
[info@eNeura.com](mailto:info@eNeura.com)  
[customercare@eNeura.com](mailto:customercare@eNeura.com)

eNeura (UK) Ltd.  
6th Floor  
One London Wall London EC2Y 5EB  
United Kingdom  
Tel. +44 (0) 20.3356 4889  
Fax: +44 (0) 20.7785.8152  
[www.eNeura.co.uk](http://www.eNeura.co.uk)  
[customercare@eneura.co.uk](mailto:customercare@eneura.co.uk)

### Authorized Representative

Emergo Europe BV  
Prinsessegracht 20  
2514 AP The Hague  
The Netherlands  
Tel: +31 (0) 70.345.8570  
Fax: +31 (0) 70.346.7299  
[www.emergoeurope.eu](http://www.emergoeurope.eu)

### Australia Sponsor

Emergo Australia  
Level 20, Tower II, Darling Park  
201 Sussex Street  
Sydney, NSW 2000  
Australia  
[www.emergogroup.com](http://www.emergogroup.com)

# SAVI™

## eNeura

© 2022 eNeura Inc.

SAVI™ is a trademarks of Aruene Corporation

The device and the method of use that comprise the eNeura SAVI are protected by U.S. Patents 6,402,678, 7,294,101, 7,494,458 and 7,601,116; by European Patents 1 307 260 B1 and 1 977 787 B1; and by Canadian Patent 2423840.

Additional patents are pending.

LBL-0188 Rev J 02/2022